

Probe Thermometer Checking Record

Probe thermometer details

Date												
Reading in Iced Water												
Reading in Boiling Water												

NB. The readings in iced water should be -1°C to +1°C, if outside this range the unit should be returned to the manufacturer.

The reading in boiling water should be between 99°C and 101°C, if outside this range the unit should be returned to the manufacturer.

These readings should be checked on a monthly basis.

Date and Details of Yearly Calibration		Date of next Yearly Calibration	
---	--	--	--

NB. The electronic display unit should be checked at least once per year.

Manufacturers may offer a service or 'calibration caps' that plug into the unit.